

Line#	XFX System, Quantity One (1)	AWARD			*See explanation	
		Bruker AXS, Inc.	Kratos Analytical	Rigaku Americas Corporation	Rigaku Americas Corporation	PANalytical
1	XRF System	\$ 192,064.70	\$ 196,590.00	\$ 171,900.00	\$ 86,300.00	\$ 214,440.00
	Make & Model	Bruker AXS S4 Pioneer	Shimadzu XRF1800	Rigaku ZSX Primus WDXRF	Rigaku ZSXmini II 3210C3	PANalytical AX10S
	Comments			discounted price	discounted price	
2	Less Trade-In	\$ (66,164.70)	\$ (5,000.00)	\$ (3,000.00)	\$ (3,000.00)	\$ (39,440.00)
Total		\$ 125,900.00	\$ 191,590.00	\$ 168,900.00	\$ 83,300.00	\$ 175,000.00
Payment Terms		0% 30	0% 0	100% 30	100% 30	
Delivery Days ARO		120	90	60	60	

***Explanation**

- 1) The equipment bid is a benchtop system. We specified that it shall be in a temperature controlled sealed cabinet equipped with castors to allow movement.
- 2) The X-ray tube does not meet the requirements requested. We specified it must be constructed with a rhodium anode of 3.0 kilowatt.
- 3) Does not meet output voltage requirements. We specify it must meet at least 60kV with an output current of 120mA.
- 4) Equipment bid is not equipped with three primary collimators: Fine, Medium and Coarse, and is not equipped with the specified three beam filters.
- 5) Does not meet the minimum of 8 crystal positions required.
- 6) Doesn't provide a dust collection device.

No Bids

- ChromTech
- GE Healthcare
- Troxler Electronic Laboratories
- VWR International
- Ward's Natural Science Establishment