

AWARD			
Bending Beam Rheometer (Qty 1)	Cannon Instrument Company	Gilson Co. Inc.	Rainhart Co.
	\$27,964.82	\$24,750.00	\$22,548.00
Payment Terms		N/A	0%,30
Delivery Days ARO	45	28-35	
Comment		Gilson HM-59 w/installation. 1C. HM-59 has mechanical refrigeration unit only advantage of the Peltier is quicker warmup of bath. Disadvantage of Peltier is larger circulator. 2C. Much easier to achieve a tighter tolerance using the precision ball bearing configuration. 1 day for installation. 1 day for Training.	

Award Explanation

Gilson Co. Inc. - Bid did not meet all of the required specifications
Rainhart Co. - Vendor did not return the spec sheet with the yes, no, other questions.

No Bids

ATS RheoSystems
Buck Scientific
ELE International