

**Attachment A 6563 OF- CORE LIST**

**Bidders must provide pricing on this bid sheet.** Attached documents will not be accepted. Bidders must bid on all items on the bid sheet **or the bid will be rejected**. Bidders must submit pricing for the units of measure on the bid sheet (e.g. kit, pkg, ea). All inclusive pricing will not be accepted. **The bidder must segregate to match the required bid format.**

Number	Estimated Quantity	Item	Model Number	DESCRIPTION	UOM	Unit Price BID
1	800	LogTag VFC-400 Data Logger Kit		Log Tag VFC-400 data logger; Fridge/Freezer NIST traceable certificate of calibration Compliant to ISO/IEC 17025:2005 2 year; 3" Stainless, 1/8" OD probe with 5' sensor cable P/N SSP-0615; High density plastic (shatterproof) glycol buffer bottle P/N SPB-1019; Acrylic stand P/N AS-1031; Extra replaceable battery; Extra replaceable battery; Wall mount bracket P/N WM-1212; Installation & Mounting Kit (Includes 5 zip ties and 3 cable tie mounts) P/N IMK-1015; Calibration and labelling; Webinar training, LogTag library, software preconfigured per customer specifications and toll free number for technical support.	KIT	
2	250	LogTag Docking Station Kit		LogTag Docking Station	PKG	
3	250	Calibration Service for LogTag Data Logger		Perform manufacturer suggested calibration as needed and/or every two (2) years.	EA	
				Each calibration must come with a new Certificate of Calibration and include the following information: Serial number Date of calibration (report or issue date) Instrument passed testing (instrument in tolerance) One of the following will be referenced within the Calibration Certificate: 1. Conforms to ISO 17025 2. Was performed by an ILAC/MRA signatory body accredited laboratory 3. Traceable to the standards maintained by NIST 4. Meets specifications and testing requirements for the American Society for Testing and Materials (ASTM) Standard E2877 tolerance Class F ( < or equal to 0.5 degrees C) or better	EA	
4	250	Inventory support service for LogTag Data Logger		Vendor will provide calibration due date tracking and notify agency of upcoming calibration deadlines per each unit	EA	